
Lab Dept: Coagulation

Test Name: RAPID TEG

General Information

Lab Order Codes: RTEG1

Synonyms: Thromboelastograph without heparinase; Rapid TEG without heparinase; RTEG

CPT Codes: 85576 – Platelet aggregation, each agent
85347 – Coagulation time, activated
85384 – Fibrinogen; activity
85390 – Fibrinolysins or coagulopathy screen, interpretation and report

Test Includes: RTEG parameters, R, K, Angle, MA, G and LY30.

Logistics

Test Indications: Assess hemorrhagic or thrombic risk by measuring rate of clot formation, strength and stability of clot; the effect of platelets, coagulation factors and cellular interactions.

Lab Testing Sections: Coagulation (Performed on Mpls campus)

Phone Numbers: MIN Lab: 612-813-6280

Test Availability: Daily, 24 hours

Turnaround Time: 2 hours

Special Instructions: Deliver immediately to the laboratory. **Specimens MUST be walked to the laboratory and cannot be delivered via pneumatic tube.** Must arrive in lab within 1 hour of collection. Indicate when specimen is drawn from a line or heparin lock.

Specimen

Specimen Type: Whole blood

Container: Light Blue top (Buffered Na Citrate 3.2%) tube

Draw Volume: 2.7 mL blood (in 3 mL tube)
Minimum: 1.8 mL blood in a 2 mL tube

Processed Volume: Same as Draw Volume

Collection: A clean venipuncture or line draw is essential. Mix specimen thoroughly by gentle inversion.

Patient's with a hematocrit level >55% must have a special tube made to adjust for the hematocrit; contact laboratory for special tube

Indicate when specimen is drawn from a line or heparin lock.

Specimens must be hand delivered to the laboratory. No pneumatic tubing of TEG specimens.

Special Processing: Lab Staff: Do Not centrifuge. Do Not freeze. Deliver to TEG workstation for immediate testing.

Patient Preparation: None

Sample Rejection: Improper tube; clotted sample; underfilled tubes; overfilled tubes; mislabeled or unlabeled tubes

Interpretive

Reference Range:

RTEG Test:	Range (all ages):
R	0.3 – 0.8 minutes
K	0.5 – 2.3 minutes
Angle	64.0 – 80.0 degrees
MA	52.0 – 71.0 mm
G	5.4 – 12.2 kd/sc
LY30	0.0 – 5.0 %

Critical Values: N/A

Limitations: Use of the pneumatic tube can alter patient TEG result values.

Methodology: TEG Viscoelastic Clot Assessment

References: Haemonetics (2014)

Haemonetics TEG 5000 Technical Manual (2008)

