

---

**Lab Dept:** Other Fluids

**Test Name:** CREATININE, FLUID

---

***General Information***

**Lab Order Codes:** FCRE (specify fluid type)

**Synonyms:** Fluid Creatinine

**CPT Codes:** 82570 – Creatinine; other source

**Test Includes:** Fluid creatinine level in mg/dL

---

***Logistics***

**Test Indications:** Useful measurement in body fluid to determine ruptured bladder.

**Lab Testing Sections:** Chemistry

**Phone Numbers:** MIN Lab: 612-813-6280

STP Lab: 651-220-6550

**Test Availability:** Daily, 24 hours

**Turnaround Time:** 1 day

**Special Instructions:** N/A

---

***Specimen***

**Specimen Type:** Body fluid (specify type)

**Container:** Green top (Li Heparin) tube

**Draw Volume:** 0.6 mL fluid

**Processed Volume:** 0.2 mL fluid, spun

**Collection:** Dependant on fluid type

**Special Processing:** N/A

**Patient Preparation:** None

**Sample Rejection:** Viscous or highly colored fluids are unacceptable

---

***Interpretive***

**Reference Range:** No reference ranges established

**Critical Values:** N/A

**Limitations:** N/A

**Methodology:** Alkaline Picrate/Kinetic

**References:** ECREA Flex® reagent cartridge insert sheet, Siemens Healthcare Diagnostics Inc, Newark, DE, 19714, PN 781270.001 – US, Issue Date 08/20/2013 RevC

EZCR Flex® reagent cartridge insert sheet (RXL/MAX), Siemens Healthcare Diagnostics Inc, Newark, DE, 19714, PN 717270.003 – US, Issue Date 11/21/2014 RevE

ECREA CAL product insert sheet, Dimension Vista System, Siemens Healthcare Diagnostics Inc, Newark, DE, 19714, PN 751270.001 – US, Issue Date 03/2013 RevD

Chemistry 1 Calibrator package insert sheet, Siemens Healthcare Diagnostics Inc, Newark, DE, 19714, PN 792018.002 – US, Issue Date 06/2013 RevE

Clinical Significance, Dade Behring Inc., Glasgow Business Community, Mailbox 531, P.O. Box 6101, Newark, Delaware 19714

Jacobs & DeMott Laboratory Test Handbook (2001) Lexi-Comp, Inc, Hudson, OH, 5th Edition